

Technical Compliance Statement



For the following information

Ref. File No.: ACI-2011349A1

Product : Field/Lab Dissolved Oxygen and Temperature Instrument

Model Number : DO210, DO200A, 9010M, 9020M

Applicant : SHANGHAI JENCO INSTRUMENTS CO., LTD

Manufacturer : SHANGHAI JENCO INSTRUMENTS CO., LTD

Standards :

EN 61326-1:2006 (IEC 61326-1:2005) (Basic immunity test requirements)
(IEC 61000-4-2:2001, IEC 61000-4-3:2006,
IEC 61000-4-4:2004+Corr.1:2006+Corr.2:2007, IEC 61000-4-6:2006)

We hereby certify that the above product has been tested by us with the listed standards and found in compliance with the council EMC directive 2004/108/EC. It is possible to use CE marking to demonstrate the compliance with this EMC Directive. The test data & results are issued on the EMC test report no. ACI-E11163A1.

Signature

Byron Kwo/Assistant General Manager

Date: Jun 07, 2012

Test Laboratory:

Audix Technology (Shanghai) Co., Ltd.

NVLAP Lab Code: 200371-0

Web Site: www.audixtech.com

NVLAP[®]
Lab Code: 200371-0

The statement is based on a single evaluation of one sample of the above-mentioned products. It does not imply an assessment of the whole production and does not permit the use of the test lab logo.